

# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm Breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

MAY 11 2009

Supervisor/Operator Performing the Verification Procedure:

DataMaster cdm S/N 130117 ✓

Name BRIAN MITCHELL

ID# 4119 Date 5-4-09 ✓

A

Agency Kodiak Police Department Phone # 907-486-8000

Instrument Location Kodiak Police Department, Booking Room

B

Alco S/N X172991 ✓ Target Value .078 ✓ High Pressure 700

Alco Test Value Average .076 ✓ .075 ✓  
1<sup>st</sup> Alco 2<sup>nd</sup> Alco

Signature

[Signature] ✓

CRS  
5/13/09

(OVER)

I, Nita J. Bolz, after being first duly sworn, depose and state as follows (to be filled in area below)

- (1) I am a Forensic Scientist IV at the State of Alaska Scientific Crime Detection Laboratory.
- (2) The Alaska Scientific Crime Detection Laboratory is an entity within the Department of Public Safety.
- (3) I am the Scientific Director of the State Breath Alcohol Program.
- (4) In that capacity, I am responsible for overseeing the breath alcohol program, including assuring that persons responsible for verifying the calibration of instruments are properly trained and qualified. I also am responsible for maintaining the records of the program.
- (5) ~~The attached verification is a true and accurate verification of calibration that reflects a regularly conducted and regularly recorded activity of the breath alcohol program performed by a person trained and qualified to conduct the verification.~~

Nita J. Bolz

Scientific Director

State Breath Alcohol Program


Subscribed and sworn before me this 22 day of May 2009.

[Signature] (Notary Seal Stamp)

Carolyn M. Noland

Notary Public, State of Alaska

Commission Expires with Office



# VERIFICATION OF CALIBRATION REPORT

of DataMaster cdm breath Test Instrument

State of Alaska

Scientific Crime Detection Laboratory - Statewide Breath Alcohol Program

DataMaster cdm S/N 130117 ✓

(CONTINUED FROM FRONT PAGE)

C

TAPE THE TEST STRIP FROM THE NONDRINKING SUBJECT TEST AND THE DIAGNOSTIC CHECK IN THE MARKED BOXES.

## NONDRINKING SUBJECT TEST

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130117

MAY 04, 2009 ✓

OPERATOR'S NAME:

MITCHELL BRIAN S

OPERATOR'S NUMBER: 4119

SUBJECT'S LAST NAME:

VERIFICATION

SUBJECT'S FIRST NAME/MI :

VERIFICATION

O.L. #: X

DEPT/AGENCY: KOD1

CASE/REPORT: X

TEST TYPE: U

ALCO TARGET VALUE: .078 ✓

ALCO S/N: K172991

### — BREATH ANALYSIS —

.078 ADJUSTED FOR 29.90 in		
ALCO TARGET	.077	00:36
BLANK TEST	.000	00:37
INTERNAL STANDARD	VERIFIED	00:37
ALCO TV 29.90 in	.076 ✓	00:38
BLANK TEST	.000	00:39
SUBJECT SAMPLE	.000	00:39
BLANK TEST	.000	00:40
ALCO TV 29.90 in	.075 ✓	00:41
BLANK TEST	.000	00:42

## DIAGNOSTIC CHECK

ALASKA DEPARTMENT OF PUBLIC SAFETY  
DATAMASTER cdm 130117

MAY 04, 2009 ✓

TIME 00:43

### — DIAGNOSTIC CHECK —

COMPUTER: OKAY

PROGRAM: OKAY

SOFTWARE DATE: 02/20/01

HEATERS

SAMPLE CHAMBER: 49c

BREATH TUBE: 43c

BAROMETER: 29.90 in

FLOW DETECTOR: OKAY

PUMP

HIGH SPEED: OKAY

DETECTOR: OKAY

FILTERS: OKAY

QUARTZ STANDARD: OKAY

CALIBRATION: OKAY

### PRINTER TEST

! " # \$ % & ' ( ) \* + , - . / 0 1 2 3 4 5 6 7 8 9 : ; < = > ? @ A B C D E F G  
H I J K L M N O P Q R S T U V W X Y Z [ \ ] ^ \_ ` a b c d e f g h i j k l m n o  
p q r s t u v w x y z ( ) \* +